

Evaluation of Oxacillin and Cefoxitin Disk and MIC Breakpoints for Prediction of Methicillin Resistance in Human and Veterinary Isolates of *Staphylococcus intermedius* Group

M. T. Wu,^{a,b} C.-A. D. Burnham,^c L. F. Westblade,^d J. Dien Bard,^e S. D. Lawhon,^f M. A. Wallace,^c T. Stanley,^g E. Burd,^{g,h} J. Hindler,^a R. M. Humphries^a

UCLA David Geffen School of Medicine, Los Angeles, California, USA^a; Landstuhl Regional Medical Center, Landstuhl, Germany^b; Washington University School of Medicine, St. Louis, Missouri, USA^c; Weill Cornell Medical College, New York, New York, USA^d; Children's Hospital of Los Angeles, Los Angeles, California, USA^e; Veterinary Medicine and Biomedical Sciences, Texas A&M University, College Station, Texas, USA^f; Emory University School of Medicine, Atlanta, Georgia, USA^g; Emory Antibiotic Resistance Center, Atlanta, Georgia, USA^h

Staphylococcus pseudintermedius is a coagulase-positive species that colonizes the nares and anal mucosa of healthy dogs and cats. Human infections with *S. pseudintermedius* range in severity from bite wounds and rhinosinusitis to endocarditis; historically, these infections were thought to be uncommon, but new laboratory methods suggest that their true incidence is underreported. Oxacillin and cefoxitin disk and MIC tests were evaluated for the detection of *mecA*- or *mecC*-mediated methicillin resistance in 115 human and animal isolates of the *Staphylococcus intermedius* group (SIG), including 111 *Staphylococcus pseudintermedius* and 4 *Staphylococcus delphini* isolates, 37 of which were *mecA* positive. The disk and MIC breakpoints evaluated included the Clinical and Laboratory Standards Institute (CLSI) M100-S25 *Staphylococcus aureus*/*Staphylococcus lugdunensis* oxacillin MIC breakpoints and cefoxitin disk and MIC breakpoints, the CLSI M100-S25 coagulase-negative *Staphylococcus* (CoNS) oxacillin MIC breakpoint and cefoxitin disk breakpoint, the CLSI VET01-S2 *S. pseudintermedius* oxacillin MIC and disk breakpoints, and the European Committee on Antimicrobial Susceptibility Testing (EUCAST) *S. pseudintermedius* cefoxitin disk breakpoint. The oxacillin results interpreted by the VET01-S2 (disk and MIC) and M100-S25 CoNS (MIC) breakpoints agreed with the results of *mecA/mecC* PCR for all isolates, with the exception of one false-resistant result (1.3% of *mecA/mecC* PCR-negative isolates). In contrast, cefoxitin tests performed poorly, ranging from 3 to 89% false susceptibility (very major errors) and 0 to 48% false resistance (major errors). BD Phoenix, bioMérieux Vitek 2, and Beckman Coulter MicroScan commercial automated susceptibility test panel oxacillin MIC results were also evaluated and demonstrated >95% categorical agreement with *mecA/mecC* PCR results if interpreted by using the M100-S25 CoNS breakpoint. The Alere penicillin-binding protein 2a test accurately detected all *mecA*-positive isolates, although for four isolates, cefoxitin induction was required prior to testing. These data demonstrate that the cefoxitin surrogate test does not reliably detect the presence of *mecA* in *S. pseudintermedius* isolates and that laboratories should perform oxacillin disk or MIC tests of these isolates when they are encountered.

Staphylococcus pseudintermedius is a coagulase-positive, hemolytic species that colonizes the nares and anal mucosa of healthy dogs and cats (1–4). Clinically, *S. pseudintermedius* is the most common cause of canine pyoderma and occasionally causes urinary tract and joint infections in dogs and cats (5–8). Phenotypically, *S. pseudintermedius* is difficult to differentiate from *Staphylococcus intermedius* and *Staphylococcus delphini*, which are also coagulase-positive veterinary staphylococci. Collectively, these three species are referred to as the *S. intermedius* group (SIG). Differentiation of the members of this group requires molecular analysis (9), although matrix-assisted laser desorption ionization-time of flight mass spectrometry (MALDI-TOF MS) has also shown promise for reliable differentiation of the members of the SIG (4, 10–13).

Human infections caused by *S. pseudintermedius* are reported sporadically and range in severity from bite wounds and rhinosinusitis to endocarditis, (14–17). However, infections caused by *S. pseudintermedius* in humans are likely underreported (18), as clinical laboratories may inaccurately report these isolates as *Staphylococcus aureus* on the basis of positive coagulase or staphylococcal latex agglutination reactions (19). Recognition of phenotypic traits atypical of *S. aureus*, such as off-white colony color and the presence of double-zone beta-hemolysis on sheep blood agar

(SBA) with appropriate follow-up testing, is required for accurate identification. The introduction of MALDI-TOF MS systems in many clinical laboratories for identification of bacteria may result in increased reporting of *S. pseudintermedius*, as has been noted by our laboratories (R.M.H. and C.-A.D.B., personal observations).

Unlike human isolates of *S. aureus*, veterinary isolates of SIG are typically susceptible to methicillin (20). However, the incidence of methicillin-resistant *S. pseudintermedius* (MRSP) in dogs has increased, presenting a major treatment challenge for veterinarians (21). In face of this emerging resistance, the Clinical and

Received 25 October 2015 Returned for modification 12 November 2015

Accepted 19 November 2015

Accepted manuscript posted online 25 November 2015

Citation Wu MT, Burnham C-AD, Westblade LF, Dien Bard J, Lawhon SD, Wallace MA, Stanley T, Burd E, Hindler J, Humphries RM. 2016. Evaluation of oxacillin and cefoxitin disk and MIC breakpoints for prediction of methicillin resistance in human and veterinary isolates of *Staphylococcus intermedius* group. J Clin Microbiol 54:535–542. doi:10.1128/JCM.02864-15.

Editor: S. S. Richter

Address correspondence to R. M. Humphries, rhumphries@mednet.ucla.edu.

Copyright © 2016, American Society for Microbiology. All Rights Reserved.

TABLE 1 Breakpoints evaluated for prediction of *S. pseudintermedius* methicillin resistance

Breakpoint standard	Oxacillin breakpoint				Cefoxitin breakpoint			
	Disk (mm)		MIC ($\mu\text{g/ml}$)		Disk (mm)		MIC ($\mu\text{g/ml}$)	
	S ^a	R ^b	S	R	S	R	S	R
CLSI M100-S25 <i>S. aureus</i> / <i>S. lugdunensis</i>	NA ^c	NA	≤ 2	≥ 4	≥ 22	≤ 21	≤ 4	≥ 8
CLSI M100-S25 CoNS	NA	NA	≤ 0.25	≥ 0.5	≥ 25	≤ 24	NA	NA
CLSI VET01-S2 <i>S. pseudintermedius</i>	≥ 18	≤ 17	≤ 0.25	≥ 0.5	NA	NA	NA	NA
EUCAST <i>S. pseudintermedius</i>	NA	NA	NA	NA	≥ 35	< 35	NA	NA

^a S, susceptible.

^b R, resistant.

^c NA, not applicable.

Laboratory Standards Institute (CLSI) veterinary group revised recommendations for the detection of methicillin resistance in these isolates (22–24) and now stipulates that oxacillin disk or MIC testing be performed and interpreted by using *S. pseudintermedius*-specific breakpoints (Table 1). These guidelines are found in the CLSI VET01-S2 document (25).

With the close contact between humans and companion animals, zoonotic transfer of MRSP to humans is inevitable. Few, if any, clinical laboratories in the United States routinely access the CLSI VET01-S2 document and rely on the CLSI M100-S25 standard for instructions on how to perform susceptibility tests and interpret their results for human isolates. This fact, coupled with frequent misidentification of *S. pseudintermedius* as *S. aureus*, has resulted in variable *S. pseudintermedius* isolate antimicrobial susceptibility testing and interpretation practices across clinical laboratories. Moreover, no automated susceptibility test system is cleared by the U.S. Food and Drug Administration (FDA) for the testing of human isolates of *S. pseudintermedius*, and the ability of these systems to accurately detect MRSP is unknown.

The objective of this study was to evaluate cefoxitin and oxacillin disk zones and MICs, compared to *mecA* and *mecC* PCR, for a collection of 115 human and veterinary SIG isolates. Disk zones and MICs were interpreted by using the VET01-S2 oxacillin breakpoints, the European Committee on Antimicrobial Susceptibility Testing (EUCAST) cefoxitin breakpoints for *S. pseudintermedius* (26), and the existing *S. aureus* and coagulase-negative *Staphylococcus* (CoNS) oxacillin and cefoxitin breakpoints in M100-S25. The results of these data were presented to the CLSI Antimicrobial Susceptibility Testing Subcommittee in July 2015, which led to the addition of *S. pseudintermedius*-specific interpretative criteria for oxacillin disk and MIC data to the forthcoming M100-S26 document. In addition, three commercial automated susceptibility test systems used in human clinical laboratories were evaluated, along with a penicillin-binding protein 2a (PBP2a) test, for the detection of MRSP.

MATERIALS AND METHODS

Bacterial isolates. A total of 115 isolates were included in this study. Human isolates ($n = 45$) were collected from the following geographically distinct academic medical centers: the Ronald Reagan UCLA Medical Center, Los Angeles, CA ($n = 12$); the Emory University School of Medicine, Atlanta, GA ($n = 4$); Washington University, St. Louis, MO ($n = 12$); the Centers for Disease Control and Prevention, Atlanta, GA ($n = 7$); the University of Iowa Hospital and Clinics, Iowa City, IA ($n = 2$); and the

R. M. Alden Research Laboratory, Culver City, CA ($n = 8$). Veterinary isolates ($n = 70$) were obtained from the Texas A&M School of Veterinary Medicine. Identification of isolates was performed according to the standard operating procedures of the submitting institution and confirmed at a central laboratory by a nucleic acid-based assay (9).

Antimicrobial susceptibility testing. Isolates were stored at -70°C in brucella broth plus 15% glycerol (BD, Sparks, MD) and subcultured twice on SBA prior to testing. Well-isolated colonies were used to prepare a suspension equivalent to a 0.5 McFarland standard. This suspension was then used to inoculate Mueller-Hinton agar (BD) for disk diffusion (DD), broth microdilution (BMD) panels and automated susceptibility cards/panels according to the CLSI (DD and BMD) or the manufacturer's instructions. BMD was performed in cation-adjusted Mueller-Hinton broth (Difco, BD) on panels prepared at UCLA according to CLSI standards (27). Oxacillin was tested at 2-fold dilutions spanning a concentration range of 0.25 to 16 $\mu\text{g/ml}$, whereas cefoxitin was tested at a single concentration of 4 $\mu\text{g/ml}$ in duplicate. For DD tests, disks containing 1 μg oxacillin or 30 μg cefoxitin (BBL, BD) were used. Both BMD and DD tests were incubated at 35°C in ambient air for 24 h, and results were examined visually by two independent readers.

Automated antimicrobial susceptibility testing was performed with the Phoenix PMIC-8 panel (BD Diagnostics Systems, Franklin Lakes, NJ), the Vitek 2 AST-GP71 card (bioMérieux, Inc., Durham, NC), and the MicroScan PM29 panel (Beckman Coulter, Inc., Brea, CA) at the Children's Hospital of Los Angeles, UCLA, and Children's Healthcare of Atlanta, respectively, in accordance with the manufacturer's specifications. Quality control was done by testing *S. aureus* ATCC 25923 (DD), *S. aureus* ATCC 29213 (BMD), *S. aureus* ATCC BAA-976, *S. aureus* ATCC BAA-977, and *S. aureus* ATCC BAA-1026 for the Vitek 2 card; *S. aureus* ATCC 29213, *S. aureus* ATCC 43300, and *S. aureus* BAA-977 for the MicroScan panel; and *S. aureus* ATCC 29213 and *S. aureus* BAA-977 for the Phoenix panel.

***mecA* and *mecC* detection, SCC*mec* typing, and repetitive-sequence PCR.** Isolates were grown on 5% sheep blood agar (Hardy Diagnostics, Santa Maria, CA), and DNA was extracted with the BiOstic Bacteremia DNA isolation kit (Mo Bio Laboratories, Carlsbad, CA). A multiplex PCR to detect and differentiate *mecA* and *mecC* (*mecA*_{LG251}) was performed by using a modification of a previously described method (36). A PCR with primers *mecCF* (5'-GAAAAAAGGCTTAGAACGCCTC-3'), *mecCR* (5'-GAAGATCTTTCCGTTTTCAGC-3'), *mecAP4* (5'-TCCA GATTACAACCTTCACCAGG-3'), and *mecAP7* (5'-CCACTTCATATCT TGTAACG-3') was performed with Ready-To-Go PCR beads (GE Healthcare). PCR products were visualized on a 1.5% agarose gel. Known *mecA*- and *mecC*-positive MRSA strains USA300 and 72796, respectively, were used as controls.

A multiplex PCR to detect and differentiate SCC*mec* types I to V was also performed on all isolates by a method described previously (28). To

TABLE 2 Summary of study isolates

Institute (state of isolate origin)	No. of isolates	Source(s)	Specimen source(s)	No. of <i>mecA</i> ⁺ isolates	SCC <i>mec</i> type (no. of isolates)	Rep-PCR clonal lineage(s) ^a
UCLA (CA)	12	Human	Blood, ethmoid sinus drainage, wound, ear, skin, abscess	3	IV (2), V (1)	A, D, F
R. M. Alden Research Laboratory (CA)	8	Human	Blood, dog bites, cat bites, wounds	0	None detected	F
Washington University (MO)	12	Human	Wounds	1	V (1)	F
Emory University School of Medicine (GA)	4	Human	Wounds	0	Not detected	A, D, E, F
University of Iowa (IA)	2	Human	Unknown	0	Not detected	A, F
CDC ^a	7	Human	Blood, wound, endotracheal aspirate	0	Not detected	A, E, F
Texas A&M ^a	70	Canine, feline, equine, porcine	Pyoderma	33	III (9), IV (8), V (9)	A, B, C, D, E, F

^a The state from which the isolate was acquired was unavailable.

analyze the phylogenetic relatedness of the isolates used in this study, repetitive-sequence PCR (rep-PCR) was performed as previously described (29, 30). The banding patterns obtained on a virtual gel were compared to determine the similarity index. Isolates with a similarity index of $\geq 95\%$ were considered to be identical. For analysis, isolates that clustered together were given similar letter designations (A to F) to indicate similarity.

PBP2a test. The Alere PBP2a Culture Colony Test (Alere Inc., Scarborough, ME) was used to test colonies grown for 24 h according to the manufacturer's instructions for *S. aureus*. All 115 isolates were tested for noninduced and induced PBP2a expression. Colonies from the same plate used to prepare inoculums for cefoxitin and oxacillin DD, as well as BMD susceptibility testing, were used for the noninduced PBP2a test. Bacteria from the edge of the cefoxitin zone of growth inhibition were harvested and used for the induced PBP2a tests. Quality control was done with the *S. aureus* ATCC 43300 (positive control) and ATCC 25923 (negative control) strains for each new lot of PBP2a tests.

Data analysis. MIC and disk results were interpreted by using the following breakpoints: (i) CLSI VET01-S2 *S. pseudintermedius* oxacillin MIC and disk, (ii) CLSI M100-S25 *S. aureus*/*S. lugdunensis* oxacillin MIC, (iii) CLSI M100-S25 *S. aureus*/*S. lugdunensis* cefoxitin MIC and disk, (iv) CLSI M100-S25 CoNS oxacillin MIC, (v) CLSI M100-S25 CoNS cefoxitin MIC and disk, and (vi) EUCAST *S. pseudintermedius* cefoxitin disk (Table 1). Categorical agreement (CA), major errors (MEs), and very MEs (VMEs) were calculated as described by Clark et al. (31). CA was defined by using the results of the *mecA/mecC* PCR. MEs were defined as findings that isolates were resistant to oxacillin or cefoxitin but negative for *mecA* and *mecC* by PCR. The ME rate was calculated by using the number of SIG isolates negative for *mecA* and *mecC* by PCR as the denominator. VMEs were defined as findings that isolates were susceptible to oxacillin or cefoxitin but positive for *mecA* or *mecC* by PCR. The VME rate was calculated by using the number of SIG isolates positive for *mecA* and *mecC* by PCR as the denominator.

Discrepant-result resolution. Isolates with an ME or VME result compared to the gold standard (*mecA/mecC* PCR) were retested by subculturing a new culture from the frozen stock and testing it by both the method that generated the error and the reference method in parallel.

RESULTS

Molecular studies. One hundred fifteen SIG isolates were included in this study, which were further identified as 111 *S. pseudintermedius* and 4 *S. delphini* isolates. Forty-five *S. pseudintermedius* isolates were from human, 56 were from canine, 7 were from feline, and 2 were from avian specimens, and 1 was from a porcine specimen (Table 2). The *S. delphini* isolates were from avian ($n =$

1) and equine ($n = 3$) specimens. *mecA* was detected in 37/115 (32.2%) isolates, all *S. pseudintermedius*, including 4/45 (8.9%) human and 33/64 (51.6%) veterinary isolates. *mecC* was not detected in any isolate in this study.

SCC*mec* typing was performed for all isolates with a multiplex PCR assay that detects and differentiates SCC*mec* types I to V. As shown in Table 2, types III ($n = 9$), IV ($n = 10$ isolates), and V ($n = 8$ isolates) were identified and found only in *mecA*-positive isolates (i.e., no *mecA* dropout events were detected). Among the human isolates, only the traditional community-associated SCC*mec* types IV and V were identified (Table 2), whereas the veterinary isolates included types III, IV, and V, as well as seven isolates that were not typeable by this SCC*mec* assay. The isolates clustered into six groups by rep-PCR, including 22 in group A (10 human, 12 veterinary), 1 in group B (veterinary), 2 in group C (veterinary), 4 in group D (2 human, 2 veterinary), 2 in group E (1 human, 1 veterinary), and 84 in group F (32 human and 52 veterinary). There was no clear division by rep-PCR of isolates from human, avian, equine, canine, or feline specimens or by the geographic locations at which the isolates were recovered.

Cefoxitin and oxacillin DD and BMD. A clear division between oxacillin MICs was found between *mecA*-positive and -negative isolates—MICs were ≥ 0.5 $\mu\text{g/ml}$ for all *mecA*-positive isolates and ≤ 0.25 $\mu\text{g/ml}$ for all *mecA*-negative isolates but one (Fig. 1). Similarly, oxacillin zones of ≤ 17 mm were observed for all *mecA*-positive isolates, and all *mecA*-negative isolates but one had oxacillin zones of ≥ 18 mm (Fig. 2). The outlier isolate was the same for both disk and MIC tests, for which the measured oxacillin zone was 13 mm and the MIC was 1 $\mu\text{g/ml}$. These results, along with the *mecA* and *mecC* PCR results, were confirmed in duplicate for this isolate. In contrast to the oxacillin test results, no clear division in cefoxitin zones was observed between *mecA*-positive and -negative isolates (Fig. 3) and only 4/37 (10.8%) *mecA*-positive isolates grew at 4 $\mu\text{g/ml}$ cefoxitin (data not shown).

Categorical interpretation of the disk and MIC results, compared to *mecA* PCR assay results, was done by using six breakpoints, and the results are presented in Table 3. With the cefoxitin CLSI M100-S25 *S. aureus*/*S. lugdunensis* breakpoints, 28/37 *mecA*-positive isolates were interpreted as susceptible by disk (i.e., 75.7% VMEs) and 33/37 were interpreted as susceptible by MIC (89.1% VMEs, Table 3). Among the *mecA*-positive isolates for which




FIG 1 Distribution of oxacillin MICs determined by BMD for 115 isolates of *S. pseudintermedius* isolates of human and animal origins.

VMEs were observed, cefoxitin zones ranged from 22 to 38 mm (Fig. 3) and MICs were $\leq 4 \mu\text{g/ml}$. All *mecA*-negative isolates were susceptible to cefoxitin by both disk and MIC tests, yielding 0% VMEs (Table 3). When the data were interpreted by the CLSI M100-

S25 *S. aureus*/*S. lugdunensis* oxacillin MIC breakpoint, 8/37 *mecA*-positive isolates were susceptible to oxacillin by MIC (21.6% VMEs). All eight VMEs were obtained with isolates with measured oxacillin MICs of 1 to 2 $\mu\text{g/ml}$, just below or at the susceptibility




FIG 2 Distribution of oxacillin growth inhibition zone diameters for 115 isolates of *S. pseudintermedius* isolates of human and animal origins.




FIG 3 Distribution of cefoxitin growth inhibition zone diameters for 115 *S. pseudintermedius* isolates of human and animal origins.

breakpoint of ≤ 2 $\mu\text{g}/\text{ml}$ (Fig. 1). In contrast, 78/78 *mecA*-negative isolates were susceptible to oxacillin (i.e., 0% MEs) (Table 3). Repeat testing confirmed all VMEs, and no trend in SCC*mec* type or rep-PCR pattern was observed for isolates that yielded VMEs for either oxacillin or cefoxitin (data not shown).

Many laboratories routinely test both oxacillin and cefoxitin MICs for *S. aureus* isolates and report an isolate as oxacillin resistant if it is resistant to either oxacillin or cefoxitin. Therefore, we evaluated if this combined strategy could correctly identify MRSP, despite VMEs observed for oxacillin and cefoxitin MICs individually. Using the M100-S25 *S. aureus*/*S. lugdunensis* oxacillin and

cefoxitin MIC breakpoints, MICs indicating susceptibility to both oxacillin and cefoxitin were observed for 7 (18.9%) of 37 *mecA*-positive isolates.

Disk and MIC data interpreted by the CLSI M100-S25 breakpoints for CoNS yielded 11/37 *mecA*-positive isolates interpreted as susceptible by cefoxitin disk (29.7% VMEs) and 0/78 *mecA*-negative isolates interpreted as resistant (0% MEs, Table 3 and Fig. 3). There are no cefoxitin MIC breakpoints for CoNS. The M100-S25 CoNS oxacillin MIC breakpoints are the same as the CLSI VET01-S2 *S. pseudintermedius* oxacillin MIC breakpoints (Table 1). Oxacillin MICs were interpreted as indicating resistance for

TABLE 3 Phenotypic susceptibility test agreement with *mecA* PCR results for 115 *S. pseudintermedius* isolates evaluated in this study

Phenotypic test and breakpoint standard	Criterion for resistance	% CA with <i>mecA</i> PCR result	No. (%) of VMEs	No. (%) of MEs
Cefoxitin DD				
CLSI M100 S25 <i>S. aureus</i> / <i>S. lugdunensis</i>	≤ 21 mm	75.7	28 (75.7)	0 (0.0)
CLSI M100 S25 CoNS	≤ 24 mm	90.4	11 (29.7)	0 (0.0)
EUCAST <i>S. pseudintermedius</i>	< 35 mm	57.4	1 (2.7)	48 (61.5)
Cefoxitin MIC, CLSI M100 S25 <i>S. aureus</i>/<i>S. lugdunensis</i>				
	≥ 4 $\mu\text{g}/\text{ml}$	71.3	33 (89.2)	0 (0.0)
Oxacillin MIC				
CLSI M100 S25 <i>S. aureus</i> / <i>S. lugdunensis</i>	≥ 4 $\mu\text{g}/\text{ml}$	93.0	8 (21.6)	0 (0.0)
CLSI M100 S25 CoNS/CLSI VET01 S2 <i>S. pseudintermedius</i>	≥ 0.5 $\mu\text{g}/\text{ml}$	99.1	0 (0)	1 (1.3)
Oxacillin DD, CLSI VET01 S2 <i>S. pseudintermedius</i>				
PBP2a uninduced, NA ^a	Positive	96.5	4 (10.8)	0 (0.0)
PBP2a induced, NA	Positive	100	0 (0.0)	0 (0.0)

^a NA, not applicable.

TABLE 4 Performance of commercial automated systems when testing 115 isolates of *S. pseudintermedius* for oxacillin susceptibility^a

System/panel (concn range [$\mu\text{g/ml}$])	CLSI M100-S25 <i>S. aureus</i> / <i>S. lugdunensis</i> oxacillin breakpoint (R, ^b $\geq 4 \mu\text{g/ml}$)			CLSI M100-S25 CoNS oxacillin breakpoint (R, $\geq 0.5 \mu\text{g/ml}$)		
	CA (%)	No. (%) of VMEs	No. (%) of MEs	CA (%)	No. (%) of VMEs	No. (%) of MEs
BD Phoenix PMIC-8 (0.25, 0.5, 1, 2)	90.4	11 (29.7)	0 (0.0)	95.7	4 (10.8)	1 (1.3)
bioMérieux Vitek2 AST-GP71 (0.5, 1, 2) ^c	93.0	8 (21.6)	0 (0.0)	98.3	1 (2.7)	1 (1.3)
Beckman Coulter MicroScan Pos MIC 29 (0.25, 0.5, 1, 2)	95.7	5 (13.5)	0 (0.0)	99.1	0 (0.0)	1 (1.3)

^a Oxacillin MICs were interpreted by using the CLSI M100-S25 *S. aureus*/*S. lugdunensis* or CoNS breakpoints; the reference is a *mecA* PCR result. Results obtained with all three systems are currently for research use only.

^b R, resistance.

^c One isolate had growth failure (repeated twice) in the AST-GP71 panel.

37/37 *mecA*-positive isolates by these breakpoints (0% VMEs) and susceptibility for 77/78 *mecA*-negative isolates (i.e., 1.3% MEs) (Table 3 and Fig. 1). The VET01-S2 document also includes oxacillin disk breakpoints for *S. pseudintermedius*, whereas oxacillin disk testing is no longer recommended for human isolates of staphylococci. Oxacillin disk test results, interpreted by the VET01-S2 breakpoint, indicated resistance for 37/37 *mecA* isolates (0% VMEs) and susceptibility for 77/78 *mecA*-negative isolates (1.3% MEs). This ME was observed for the same *mecA*-negative isolate that was oxacillin resistant by MIC (Fig. 2).

EUCAST cefoxitin disk breakpoints for *S. pseudintermedius* (Table 1) were evaluated. One of 37 *mecA*-positive isolates was interpreted as susceptible (2.7% VMEs). This isolate had a measured cefoxitin zone of 38 mm, which was reproduced twice. In contrast, 48/78 *mecA*-negative isolates were resistant to cefoxitin by the EUCAST disk breakpoint, a 61.5% ME rate.

Performance comparison of automated susceptibility test systems. The performance of three commercial, automated susceptibility test systems, the BD Phoenix, bioMérieux Vitek 2, and Beckman Coulter MicroScan, was assessed. The test panels, MIC ranges, and performance compared to *mecA* PCR are summarized in Table 4. Notably, 0/37 *mecA*-positive isolates had a cefoxitin MIC of $>4 \mu\text{g/ml}$ —which represents a 100% cefoxitin VME rate for all three antimicrobial susceptibility test systems (not shown). Given this poor performance, only oxacillin MICs were evaluated, and they were interpreted manually by using the CLSI M100-S25 *S. aureus*/*S. lugdunensis* and M100-S25 CoNS/VET01-S2 *S. pseudintermedius* oxacillin MIC breakpoints. When interpreted by the *S. aureus*/*S. lugdunensis* oxacillin breakpoint, 11 (29.7%), 8 (21.6%), and 5 (13.5%) VMEs were obtained with the Phoenix, Vitek 2, and MicroScan systems, respectively. No MEs were observed. When evaluated by the M100-S25 CoNS/VET01-S2 *S. pseudintermedius* breakpoint, 4 (10.8%), 1 (2.7%), and 0 (0.0%) VMEs were obtained with the Phoenix, Vitek 2, and MicroScan systems, respectively. When these isolates were retested, the VME obtained with the Vitek 2 system was resolved (i.e., the MIC of 0.5 $\mu\text{g/ml}$ was $\leq 0.25 \mu\text{g/ml}$ on retesting) and all of the VMEs obtained with the Phoenix system were reproduced. Two of the Phoenix VMEs were obtained with isolates that also had lower MICs in the other commercial test systems. One, a canine isolate, had an oxacillin MIC of 0.5 $\mu\text{g/ml}$ in the Vitek 2 system and 1 $\mu\text{g/ml}$ in the MicroScan system, and the other, a porcine isolate, had oxacillin MICs of 0.5 and 1 $\mu\text{g/ml}$ in those systems, respectively. Notably, all four isolates had oxacillin MICs of $>16 \mu\text{g/ml}$ by the BMD method. One ME (1.3%) was observed when the M100-S25 CoNS/VET01-S2 *S. pseudintermedius* breakpoint was

used in all three systems, for the same canine isolate that yielded a ME by the BMD and DD methods. One growth failure was observed, for a *mecA*-positive isolate, in the Vitek 2 system. This was confirmed by repeat testing. The overall CA levels when using the M100-S25 CoNS/VET01-S2 oxacillin *S. pseudintermedius* breakpoint were 95.7, 98.3, and 99.1%, for the Phoenix, Vitek 2, and MicroScan systems, respectively.

PBP2a test. The Alere PBP2a test, when performed with colonies from the edge of the cefoxitin zone, was 100% sensitive and 100% specific, compared to the *mecA* PCR result. In contrast, if it was performed uninduced, following the package insert instructions for *S. aureus*, four *mecA*-positive isolates were repeatedly negative for PBP2a ($n = 3$ replicates), yielding a sensitivity of 89.2% and a specificity of 100%. No correlation was found between isolates that required induction and oxacillin MIC or cefoxitin disk zones (data not shown).

DISCUSSION

Clinical laboratories almost universally rely on the use of the surrogate agent cefoxitin to detect *mecA*-mediated beta-lactam resistance when testing human isolates of staphylococci (27). The cefoxitin disk test has been shown in several studies to be the most reliable predictor of the presence of *mecA* in both *S. aureus* and CoNS. However, the present study demonstrates that this test falls short for SIG isolates of both human and veterinary origins. Regardless of the breakpoint employed, cefoxitin VME rates compared to *mecA* PCR were unacceptably high and ranged from 3 to 89% (Table 3). On the low end of this range was the EUCAST cefoxitin disk test, with 3% VMEs observed among the 37 *mecA*-positive isolates tested. However, we obtained a very high ME rate (62.8%) when using this breakpoint, further demonstrating the limitation of this surrogate agent for SIG. Indeed, we were unable to find a clear breakpoint in cefoxitin zones that could differentiate *mecA*-positive from *mecA*-negative isolates (Fig. 3). No correlation between SCC*mec* type and/or rep-PCR clonal type and performance of cefoxitin tests was identified, suggesting that our findings were not related to a specific lineage of this organism group. It should be noted that we incubated all tests for 24 h, which is longer than the recommended 16- to 20-h incubation endorsed by EUCAST (26)—we cannot exclude the possibility that the additional 4 h of incubation contributed to some MEs. Nonetheless, it is clear that neither the cefoxitin nor the oxacillin test will adequately detect *mecA*-mediated resistance in all species of *Staphylococcus*—both VMEs and MEs (32) have been documented. In particular, the cefoxitin disk test performs poorly with *S. saprophyticus* (32) and *S. simulans* (33) when results are inter-

preted by using the CLSI CoNS breakpoints. The mechanisms behind these discrepancies are unclear, but the mechanism of beta-lactam resistance in the staphylococci is complex, involving the participation of both acquired and native PBP-encoding genes; therefore, differences in phenotypic resistance among the various staphylococcal species are not unanticipated (34).

Oxacillin MIC and disk testing, when interpreted by the VET01-S2 *S. pseudintermedius* breakpoints, was most accurate at identifying *mecA* among the SIG isolates evaluated. All 37 *mecA*-positive isolates were correctly interpreted as oxacillin resistant, and 77/78 *mecA/mecC*-negative isolates were interpreted as oxacillin susceptible. Similar to our findings, Bemis and colleagues found 29.1% VMEs and 0% MEs for the cefoxitin disk test among a collection of 380 veterinary isolates of *S. pseudintermedius* (88 *mecA* positive) (22). This laboratory also tested *S. pseudintermedius* with the oxacillin disk test interpreted by the CLSI historical CoNS oxacillin breakpoint that was last published in 2008 in CLSI M100-S18 (and is the same as the current CLSI VET01-S2 *S. pseudintermedius* disk breakpoint). Among 666 isolates (230 *mecA* positive), they found 0.9% VMEs and 0.6% MEs. Similarly, when Schissler and colleagues evaluated a collection of 30 *mecA*-positive veterinary isolates of *S. pseudintermedius*, only 6.7 or 43.3% were resistant when the cefoxitin disk test was used and interpreted according to the CLSI M100-S25 *S. aureus/S. lugdunensis* or CoNS breakpoint, respectively. These findings, which correlated with our study of human isolates, confirm that cefoxitin is not predictive of *mecA*-mediated resistance in *S. pseudintermedius* isolates, regardless of the isolate source. In response to the data presented here, the CLSI Antimicrobial Susceptibility Testing Subcommittee voted in June 2015 to add *S. pseudintermedius*-specific oxacillin MIC and disk breakpoints to the forthcoming M100-S26 document. These new breakpoints for human isolates of *S. pseudintermedius* are the same as those published in the VET01-S2 document and will help raise awareness regarding the use of oxacillin tests to detect *mecA*-mediated resistance in *S. pseudintermedius* among clinical laboratories, in particular as cefoxitin testing has been recommended and oxacillin testing has been discouraged by the CLSI for other species of staphylococci. However, the use of these new oxacillin breakpoints, when they are published, is contingent on the laboratory's accurate identification of *S. pseudintermedius* when it is isolated from diagnostic specimens. Accurate identification may be a challenge to laboratories that do not have MALDI-TOF MS technology for bacterial identification and impossible for laboratories that rely solely on a coagulase or *S. aureus* latex test for the identification of staphylococci, as both tests yield positive reactions with SIG isolates. We have found that commercial systems will identify isolates of *S. pseudintermedius* as *S. intermedius* (L. Westblade, unpublished data). Thus, if these isolates are encountered, laboratories should perform oxacillin MIC or disk tests on them to predict the presence of *mecA*.

Nonetheless, the inclusion of oxacillin disk breakpoints in M100-S26 is particularly helpful to laboratories in the United States, as no commercial antimicrobial susceptibility test systems are cleared by the FDA for *S. pseudintermedius*. While *S. pseudintermedius* is the only *Staphylococcus* species for which the oxacillin disk test would be used, many laboratories use oxacillin DD as a surrogate test for penicillin susceptibility in isolates of *Streptococcus pneumoniae*. These laboratories would thus already have the disk on hand and quality controlled. As an alternative, the commercial antimicrobial susceptibility test systems performed well;

oxacillin MICs interpreted by M100-S25 CoNS breakpoints from all three instruments demonstrated a >95% CA. These systems appear to be able to detect *mecA*-mediated resistance if the CoNS breakpoints are used, with the exception of the BD Phoenix system, where a 10.8% VME rate was noted. Importantly, 0 of the 37 *mecA*-positive isolates had a positive cefoxitin screening result in these systems, reinforcing the poor performance of this surrogate agent for SIG and the potential negative implication of misidentification of SIG as *S. aureus*. Furthermore, it is worth noting that we encountered some difficulty extracting oxacillin results from the commercial platforms. For example, the Phoenix system required the operator to annotate "unspecified" for an organism's identification in order to retrieve oxacillin MIC results, and Vitek 2 required similar manipulation of the platform's software and/or setting to obtain results. Laboratories should be aware that while the results of the present study are encouraging, they are reported for research use only and would require laboratory verification prior to use in determining patient care. Therefore, the oxacillin disk test is an attractive option.

The Alere PBP2a Culture Colony test demonstrated 100% agreement with the *mecA* PCR results, if colonies from the edge of the cefoxitin zone were used for testing. Cefoxitin, rather than oxacillin, was used as the inducer for *mecA* expression in our study on the basis of literature about *S. aureus* that demonstrates that cefoxitin is the superior inducer of *mecA* expression in this species (35). It is perhaps fortuitous that we chose this approach, since it allows laboratories that do not stock an oxacillin disk to perform a PBP2a test for *S. pseudintermedius* with colonies from the edge of the cefoxitin disk (which is stocked by nearly all laboratories in the United States) zone of inhibition.

Throughout this study, we encountered one isolate, of canine origin, that yielded a consistently elevated oxacillin MIC and small oxacillin zones but was *mecA* and *mecC* negative by PCR. This isolate had an oxacillin zone diameter of 12 mm, a cefoxitin zone diameter of 27 mm, and an oxacillin MIC of >16 µg/ml but a cefoxitin MIC of ≤4 µg/ml. The oxacillin MICs were similarly high (1 to >2 µg/ml) by all three commercial systems. The mechanism of oxacillin resistance in this isolate remains to be determined, but it may be due to either hyperexpression of *blaZ* or other altered penicillin-binding proteins not detected by our molecular analysis-based methods. While these alternative mechanisms of low-level oxacillin resistance have been documented in *S. aureus*, they have not, to our knowledge, been confirmed in isolates of SIG.

In summary, our data demonstrate that the VET01-S2 oxacillin disk and MIC breakpoints are appropriate for use with *S. pseudintermedius* isolates obtained from human clinical specimens. Laboratories should attempt to identify coagulase-positive staphylococci that appear atypical in morphology (double zone of hemolysis, white in color) by MALDI-TOF MS or nucleic acid methods to ensure the appropriate determination and interpretation of oxacillin MICs and DD zones.

ACKNOWLEDGMENTS

We thank Brandi Limbago and Kamille Rasheed of the Centers for Disease Control and Prevention and Bradley Ford of the University of Iowa for providing isolates included in this study. We also thank Paul Magnano, Anita Sokovic, Karina Hernandez, and Marissa Carvalho for their invaluable technical support throughout this study. We thank Alere for providing the PBP2a tests used in this study and BD, Beckman Coulter,

and bioMérieux for providing antimicrobial susceptibility test panels for this study.

REFERENCES

- Cox HU, Hoskins JD, Newman SS, Foil CS, Turnwald GH, Roy AF. 1988. Temporal study of staphylococcal species on healthy dogs. *Am J Vet Res* 49:747–751.
- Cox HU, Hoskins JD, Newman SS, Turnwald GH, Foil CS, Roy AF, Kearney MT. 1985. Distribution of staphylococcal species on clinically healthy cats. *Am J Vet Res* 46:1824–1828.
- Rubin JE, Ball KR, Chirino-Trejo M. 2011. Antimicrobial susceptibility of *Staphylococcus aureus* and *Staphylococcus pseudintermedius* isolated from various animals. *Can Vet J* 52:153–157.
- Sasaki T, Kikuchi K, Tanaka Y, Takahashi N, Kamata S, Hiramatsu K. 2007. Methicillin-resistant *Staphylococcus pseudintermedius* in a veterinary teaching hospital. *J Clin Microbiol* 45:1118–1125. <http://dx.doi.org/10.1128/JCM.02193-06>.
- Rubin JE, Gaunt MC. 2011. Urinary tract infection caused by methicillin-resistant *Staphylococcus pseudintermedius* in a dog. *Can Vet J* 52:162–164.
- Nienhoff U, Kadlec K, Chaberny IF, Verspohl J, Gerlach GF, Schwarz S, Kreienbrock L, Nolte I, Simon D. 2011. Methicillin-resistant *Staphylococcus pseudintermedius* among cats admitted to a veterinary teaching hospital. *Vet Microbiol* 153:414–416. <http://dx.doi.org/10.1016/j.vetmic.2011.05.045>.
- Weese JS, Faires MC, Frank LA, Reynolds LM, Battisti A. 2012. Factors associated with methicillin-resistant versus methicillin-susceptible *Staphylococcus pseudintermedius* infection in dogs. *J Am Vet Med Assoc* 240:1450–1455. <http://dx.doi.org/10.2460/javma.240.12.1450>.
- Beck KM, Waisglass SE, Dick HL, Weese JS. 2012. Prevalence of methicillin-resistant *Staphylococcus pseudintermedius* (MRSP) from skin and carriage sites of dogs after treatment of their methicillin-resistant or methicillin-sensitive staphylococcal pyoderma. *Vet Dermatol* 23:369–375. <http://dx.doi.org/10.1111/j.1365-3164.2012.01035.x>.
- Sasaki T, Tsubakishita S, Tanaka Y, Sakusabe A, Ohtsuka M, Hirota S, Kawakami T, Fukata T, Hiramatsu K. 2010. Multiplex-PCR method for species identification of coagulase-positive staphylococci. *J Clin Microbiol* 48:765–769. <http://dx.doi.org/10.1128/JCM.01232-09>.
- Devriese LA, Vancannet M, Baele M, Vanechoutte M, De Graef E, Snauwaert C, Cleenwerck I, Dawyndt P, Swings J, Decostere A, Haesebrouck F. 2005. *Staphylococcus pseudintermedius* sp. nov., a coagulase-positive species from animals. *Int J Syst Evol Microbiol* 55:1569–1573. <http://dx.doi.org/10.1099/ijs.0.63413-0>.
- Silva MB, Ferreira FA, Garcia LN, Silva-Carvalho MC, Botelho LA, Figueiredo AM, Vieira-da-Motta O. 2015. An evaluation of matrix-assisted laser desorption ionization time-of-flight mass spectrometry for the identification of *Staphylococcus pseudintermedius* isolates from canine infections. *J Vet Diagn Invest* 27:231. <http://dx.doi.org/10.1177/1040638715573297>.
- Chrobak D, Kizerwetter-Swida M, Rzewuska M, Binek M. 2011. Antibiotic resistance of canine *Staphylococcus intermedium* group (SIG)—practical implications. *Pol J Vet Sci* 14:213–218.
- Sasaki T, Kikuchi K, Tanaka Y, Takahashi N, Kamata S, Hiramatsu K. 2007. Reclassification of phenotypically identified *Staphylococcus intermedium* strains. *J Clin Microbiol* 45:2770–2778. <http://dx.doi.org/10.1128/JCM.00360-07>.
- Starlander G, Börjesson S, Gronlund-Andersson U, Tellgren-Roth C, Melhus A. 2014. Cluster of infections caused by methicillin-resistant *Staphylococcus pseudintermedius* in humans in a tertiary hospital. *J Clin Microbiol* 52:3118–3120. <http://dx.doi.org/10.1128/JCM.00703-14>.
- Riegel P, Jesel-Morel L, Laventie B, Boisset S, Vandenesch F, Prevost G. 2011. Coagulase-positive *Staphylococcus pseudintermedius* from animals causing human endocarditis. *Int J Med Microbiol* 301:237–239. <http://dx.doi.org/10.1016/j.ijmm.2010.09.001>.
- Stegmann R, Burnens A, Maranta CA, Perreten V. 2010. Human infection associated with methicillin-resistant *Staphylococcus pseudintermedius* ST71. *J Antimicrob Chemother* 65:2047–2048. <http://dx.doi.org/10.1093/jac/dkq241>.
- Van Hoovels L, Vankeerberghen A, Boel A, Van Vaerenbergh K, De Beenhouwer H. 2006. First case of *Staphylococcus pseudintermedius* infection in a human. *J Clin Microbiol* 44:4609–4612. <http://dx.doi.org/10.1128/JCM.01308-06>.
- Börjesson S, Gomez-Sanz E, Ekstrom K, Torres C, Gronlund U. 2015. *Staphylococcus pseudintermedius* can be misdiagnosed as *Staphylococcus aureus* in humans with dog bite wounds. *Eur J Clin Microbiol Infect Dis* 34:839–844. <http://dx.doi.org/10.1007/s10096-014-2300-y>.
- Pottumarthy S, Schapiro JM, Prentice JL, Houze YB, Swanzy SR, Fang FC, Cookson BT. 2004. Clinical isolates of *Staphylococcus intermedium* masquerading as methicillin-resistant *Staphylococcus aureus*. *J Clin Microbiol* 42:5881–5884. <http://dx.doi.org/10.1128/JCM.42.12.5881-5884.2004>.
- van Duijkeren E, Box AT, Heck ME, Wannet WJ, Fluit AC. 2004. Methicillin-resistant staphylococci isolated from animals. *Vet Microbiol* 103:91–97. <http://dx.doi.org/10.1016/j.vetmic.2004.07.014>.
- Gortel K, Campbell KL, Kakoma I, Whittam T, Schaeffer DJ, Weisiger RM. 1999. Methicillin resistance among staphylococci isolated from dogs. *Am J Vet Res* 60:1526–1530.
- Bemis DA, Jones RD, Frank LA, Kania SA. 2009. Evaluation of susceptibility test breakpoints used to predict *mecA*-mediated resistance in *Staphylococcus pseudintermedius* isolated from dogs. *J Vet Diagn Invest* 21:53–58. <http://dx.doi.org/10.1177/104063870902100108>.
- Frank LA, Loeffler A. 2012. Methicillin-resistant *Staphylococcus pseudintermedius*: clinical challenge and treatment options. *Vet Dermatol* 23:283–291. <http://dx.doi.org/10.1111/j.1365-3164.2012.01047.x>.
- Schissler JR, Hillier A, Daniels JB, Cole LK, Gebreyes WA. 2009. Evaluation of Clinical and Laboratory Standards Institute interpretive criteria for methicillin-resistant *Staphylococcus pseudintermedius* isolated from dogs. *J Vet Diagn Invest* 21:684–688. <http://dx.doi.org/10.1177/104063870902100514>.
- CLSI. 2013. Performance standards for antimicrobial disk and dilution susceptibility tests for bacteria isolated from animals VET01–S2. Second information supplement. Clinical and Laboratory Standards Institute, Wayne, PA.
- EUCAST. 2015. *Staphylococcus pseudintermedius* cefoxitin 30 µg as screen for methicillin resistance, version 1.1. European Committee on Antimicrobial Susceptibility Testing, Basel, Switzerland.
- CLSI. 2015. Performance standards for antimicrobial susceptibility testing, 25th informational supplement. M100–S25. Clinical and Laboratory Standards Institute, Wayne, PA.
- Boye K, Bartels MD, Andersen IS, Moller JA, Westh H. 2007. A new multiplex PCR for easy screening of methicillin-resistant *Staphylococcus aureus* SCC*mec* types I–V. *Clin Microbiol Infect* 13:725–727. <http://dx.doi.org/10.1111/j.1469-0691.2007.01720.x>.
- El Feghaly RE, Stamm JE, Fritz SA, Burnham CA. 2012. Presence of the *bla*(Z) beta-lactamase gene in isolates of *Staphylococcus aureus* that appear penicillin susceptible by conventional phenotypic methods. *Diagn Microbiol Infect Dis* 74:388–393. <http://dx.doi.org/10.1016/j.diagmicrobio.2012.07.013>.
- Frye SHM. 2006. Molecular strain typing using repetitive sequence-based PCR, p 540. In Tang Y-W, Stratton CW (ed), *Advanced techniques in diagnostic microbiology*. Springer, New York, NY.
- Clark RB LM, Loeffelholz MJ, Thibbetts RJ. 2009. Cumitech 31A, Verification and validation of procedures in the clinical microbiology laboratory. ASM Press, Washington, DC.
- Johnson KN, Andreatchio K, Edelstein PH. 2014. Detection of methicillin-resistant coagulase-negative staphylococci by the Vitek 2 system. *J Clin Microbiol* 52:3196–3199. <http://dx.doi.org/10.1128/JCM.01162-14>.
- Swenson JM, Tenover FC. 2005. Results of disk diffusion testing with cefoxitin correlate with presence of *mecA* in *Staphylococcus* spp. *J Clin Microbiol* 43:3818–3823. <http://dx.doi.org/10.1128/JCM.43.8.3818-3823.2005>.
- da Costa Darini AL, Palazzo IC. 2004. Cefoxitin does not induce production of penicillin binding protein 2a in methicillin-susceptible *Staphylococcus aureus* strains. *J Clin Microbiol* 42:4412; author reply, 4412–4413. <http://dx.doi.org/10.1128/JCM.42.9.4412-4413.2004>.
- Boutiba-Ben Boubaker I, Ben Abbes R, Ben Abdallah H, Mamlouk K, Mahjoubi F, Kammoun A, Hammami A, Ben Redjeb S. 2004. Evaluation of a cefoxitin disk diffusion test for the routine detection of methicillin-resistant *Staphylococcus aureus*. *Clin Microbiol Infect* 10:762–765. <http://dx.doi.org/10.1111/j.1469-0691.2004.00919.x>.
- Stegger M, Andersen PS, Kearns A, Pichon B, Holmes MA, Edwards G, Laurent F, Teale C, Skov R, Larsen AR. 2012. Rapid detection, differentiation and typing of methicillin-resistant *Staphylococcus aureus* harbouring either *mecA* or the new *mecA* homologue *mecA*(LGA251). *Clin Microbiol Infect* 18:395–400. <http://dx.doi.org/10.1111/j.1469-0691.2011.03715.x>.